

1 **The Valgent4 Protocol: Robust analytical and clinical validation of 11 HPV**
2 **assays with genotyping on cervical samples collected in SurePath medium.**

3 Jesper Bonde¹, Ditte Møller Ejegod¹, Kate Cuschieri², Joakim Dillner³, Daniëlle A.M. Heideman⁴, Wim
4 Quint⁵, Miguel Angel Pavon Ribas⁶, Elizaveta Padalko⁷, Irene Kraus Christiansen⁸, Lan Xu⁹, & Marc Arbyn⁹

5 1. Department of Pathology, Copenhagen University Hospital, Kettegårds Alle 30, 2650 Hvidovre, Denmark.

6 2. Scottish HPV Reference Laboratory, Royal Infirmary of Edinburgh, EH16 4SA, Edinburgh, Scotland.

7 3. International HPV Reference Center, Karolinska University Hospital, Alfred Nobels Allé 8, Huddinge
8 Stockholm, Sweden.

9 4. Cancer Center Amsterdam, Department of Pathology, VU University Medical Center, De Boelelaan 1118,
10 1081 HV Amsterdam, The Netherlands.

11 5. DLL Diagnostic Laboratory, Visseringlaan 25, 2288 ER, Rijswijk, Netherlands.

12 6. Infection and Cancer Laboratory, Cancer Epidemiology Research Program and CIBER-ONC, Granvia de
13 L'Hospitalet 199-203, Institut Català d'Oncologia, Barcelona, Spain.

14 7. Department of Laboratory Medicine, Ghent University Hospital, Sint-Pietersnieuwstraat 25, B-9000
15 Ghent, Belgium.

16 8. Norwegian HPV Reference Laboratory, Department of Microbiology and Infection Control, Akerhus
17 University Hospital, 1478 Lørenskog, Norway.

18 9. Belgian Cancer Centre/Unit of Cancer Epidemiology, Scientific Institute of Public Health, Juliette
19 Wytsmanstreet 14, B1050 Brussels, Belgium.

20

21 Keywords: HPV genotyping assays, The VALGENT framework, cervical cancer screening, Clinical validation of
22 HPV assays

23

24 Word count: 3960

25 High lights

26

27 Abbreviations:

28 VALGENT: VALidation of HPV GENotyping Tests

29 HPV Human Papillomavirus

30 ASCUS: Atypical Squamous Cells of Undetermined Significance

31 LSIL: Low-grade squamous intraepithelial lesion

32 HSIL: High-grade squamous intraepithelial lesion

33 CIN: Cervikal intraepitelial neoplasi

34 Hr: High Risk

35 DNA: Deoxyribonucleic acid

36 IARC: International Agency for Research on Cancer

37 LBC: liquid based cytology

38 PCR: Polymerase Chain Reaction

39 EIA: Enzyme immunoassay

40 Patobank: Danish Pathology Data Bank

41

42 **Abstract**

43 Background: The VALidation of HPV GENotyping Tests (VALGENT) is an international initiative
44 designed to validate HPV assays with genotyping capability. The VALGENT4 protocol differs from
45 previous VALGENT installments as the sample collection medium is SurePath, and exclusively
46 includes samples from women ≥ 30 years of age which is concordant with the majority of HPV
47 primary screening guidelines. Here we present the protocol for the fourth installment of the
48 VALGENT framework.

49 Objectives: In VALGENT4 11 HPV assays will be evaluated using two comparator assays based on
50 PCR with the GP5+/6+ primers.

51 Study design: Overall, the VALGENT4 panel consists of 1,297 routine samples comprised of 998
52 unselected, consecutive samples, of which 51 samples had abnormal cytology with 13 women
53 diagnosed with $\geq \text{CIN}2$, and 299 consecutive samples enriched for $\geq \text{ASCUS}$ cytology (100 ASCUS,
54 100 LSIL, 99 HSIL) with 106 $\geq \text{CIN}2$ upon follow up. Manipulated and DNA extracted panel samples
55 were characterized with respect to human beta globin (HBB) and overall DNA content and
56 composition to quality assess the panel prior to distribution to the collaborating sites.

57

58 Result: The relative cellularity (mean CT value of HBB from the Onclarity assay) on the 1,297 LBC
59 samples (CT=24.8) was compared with 293 un-manipulated routine cytology screening samples
60 (CT=23.8). Furthermore, the DNA extracted panel samples was characterized using the Exome
61 iPLEX pro assay, which reports amplifiable copies on individual samples as well as copies of five
62 different base pair lengths. Here the data showed a slightly lower number of amplifiable DNA
63 copies (ratio: 0.7, $p < 0.01$) in the VALGENT4 panel samples compared to routine extracted
64 cervical DNA samples

65 Conclusion: The present manuscript details the manipulation, processing and quality assessment
66 of samples used in VALGENT-4. This methodological document may be of value for future
67 international projects of HPV test validation.

68 **Background**

69 Practically all cervical cancers are derived from an Human papillomavirus (HPV) infection [1-3] and the causal
70 relation is firmly established between the development of cervical cancer and at least 12 HPV genotypes (HPV
71 16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, and HPV59, all group 1 carcinogens as classified by IARC) [1, 2].
72 Evidence on the clinical value of HPV testing for triage of borderline cervical cytology and test of cure in
73 women who have been treated for pre-invasive cervical lesions is today widely accepted [4-6]. Additionally,
74 randomized controlled trials have provided evidence for the use of HPV-based screening over cytology as a
75 superior and highly reproducible screening modality to prevent both invasive squamous and adeno-
76 carcinomas [5, 7]. As a result, primary HPV-based cervical screening is being implemented in several countries
77 globally, substituting conventional and liquid based cytology (LBC) as the standard of care. Together, this has
78 created a rapidly expanding market for molecular HPV tests, and more than 200 different HPV assays are
79 now commercially available [8]. As primary HPV based screening becomes the standard of care, additional

80 focus on evidence of clinical performance of HPV assays for use in cervical cancer screening programs is
81 required.

82 HPV assays, applicable to cervical screening and disease management contexts, have undergone a rapid
83 technical and scientific evolution over the last decade. The first generation of clinical HPV assays were
84 developed solely to detect oncogenic HPV genotypes using DNA PCR or hybridization techniques and mainly
85 reported the test outcome as either HPV positive or negative, with no individual HPV genotype information
86 available. Newer generations of commercially available HPV assays depend upon a wide variety of DNA and
87 RNA detection techniques and allow for individual reporting of HPV genotypes. Current HPV genotyping
88 assays can basically be divided into three categories 1) Assays with limited genotyping that report separate
89 identification of HPV16 and HPV18 or HPV18/HPV45, combined with pooled detection of the remainder of
90 the oncogenic types, 2) Assays with extended genotyping that report separate identification of ≥ 5 genotypes
91 combined with one or more bulk detections of the remainder, and 3) Assays with full genotyping, reporting
92 individual identification of all carcinogenic HPV genotypes [9].

93 Studies have assigned different carcinogenic potential, or risk, to individual HPV genotypes, with HPV16 being
94 the most potent, in particular for squamous carcinomas of the cervix [10-15], followed by HPV18 and to some
95 extent HPV45 associated with adenocarcinomas of the cervix [12, 16-18]. Together, up to 70 % of all cervical
96 cancers are caused by HPV16 and HPV18. An additional five HPV genotypes (HPV31, 33, 45, 52, & 58) add
97 around 19 % to the burden of cervical cancer incidence [16]. Recent study data from long term follow up of
98 a Danish cervical screening cohort, have shown that persistent infections with HPV16, HPV18, HPV31 and
99 HPV33 are associated with an increased risk for development of cervical intraepithelial neoplasia grade 3 or
100 worse (\geq CIN3) compared to the remaining oncogenic genotypes [14, 19]. The six carcinogenic types HPV35,
101 39, 51, 56, 59, 68 in contrast contribute 8-9 % of the cumulative proportion of HPV-induced cervical cancer
102 [16, 20]. Other HPV genotypes incl. HPV66 and 68 are only rarely involved in cervical carcinogenesis, but both
103 are included in all oncogenic HPV assays [16]. Finally, genotype distribution among highly vaccinated cohorts

104 of women have drastically changed affecting the performance of both cytology screening and HPV
105 genotyping [21].

106 Taken together, knowledge of the genotype(s) in a screening sample may add information on the individual
107 woman's risk of developing disease [12]. Consequently, HPV technology is progressing from simple
108 plus/minus outcomes towards revealing more detailed information on the HPV infection in question. This
109 fact is reflected in a growing number of national screening guidelines where HPV16/HPV18 genotyping is
110 integrated in screening algorithms [5, 22, 23].

111 Fulfillment of the "*Guidelines for human papillomavirus DNA test requirements for primary cervical cancer*
112 *screening in women of 30 years and older*" [24], is widely accepted as the methodology by which to validate
113 HPV assays for screening use [24, 25] and several limited or extended genotyping HPV assays have fulfilled
114 the international guidelines [25-32]. The international validations are based upon the overall clinical
115 performance but not with respect to the clinical performance of detection of individual genotypes *per se*.
116 Assay specific detection of individual genotypes has instead been addressed in studies of defined plasmid
117 based panels with known copy numbers [33, 34]. The VALidation of HPV GENotyping Tests (VALGENT)
118 initiative was established to create a framework for comparison and validation of HPV genotyping assays
119 using validated comparator assays, similar to the international guidelines [24] and using a relevant sample
120 population with sufficient disease to confirm clinical performance while including primary screening samples
121 [9]. So far, three installments of VALGENT protocols have occurred, where a broad variety of HPV assays with
122 both extended and full genotyping capability have been evaluated [9].

123 **Objectives**

124 The fourth installment of the VALGENT framework, the VALGENT4 study, aims to provide high quality
125 comparative data on clinical performance of HPV assays on cervical screening samples. The VALGENT4
126 protocol differs from previous VALGENT installments as the sample collection medium is BD SurePath™
127 (Becton, Dickinson and Company), targets DNA assays only, and exclusively includes only samples from

128 women ≥ 30 years of age which is concordant with the majority of HPV primary screening guidelines. In
129 addition, a high number of assays previously untested on this LBC collection media will be evaluated.
130 Moreover, the VALGENT4 protocol is unique in the VALGENT iterations by proposing a new standard for
131 describing the quality of the sample panel included.

132 **Study Design**

133 **Study Description**

134 Sample collection, registration, processing and aliquoting was done at the parent site at the Molecular
135 Pathology Laboratory, Dept. Pathology, Copenhagen University Hospital, Hvidovre. The VALGENT4 samples
136 will be provided to participating scientific partners anonymized in concordance with Danish Data Protection
137 Agency regulations. Aliquot panels were shipped refrigerated from parent laboratory to scientific partners.

138 The VALGENT4 panel was generated using fresh SurePath collected screening samples from the Danish
139 cervical cancer screening program which services a well-screened population with a high background risk of
140 cervical cancer[35]. Moreover, the parent laboratory delivered the panel in two defined versions; pre-
141 extracted and quality described DNA for assays requiring DNA as input material (time between collection and
142 DNA extraction: Mean 28 days, min: 2 days, max: 70 days), and original LBC material (time between collection
143 and aliquoting for panel: Mean 16 days, min: 2 days, and max: 67 days) to those laboratories where assays
144 with full, integrated CE-IVD marked work flows were interrogated. The option to have a panel consisting of
145 extracted DNA limits any variability of various 3rd party DNA extraction platforms on HPV assay performance.
146 Finally, a novel element entailed characterization of all included panel samples using a MALDI-TOF assay
147 which measured the available DNA as well as the relative level of DNA fragmentation in the individual sample.

148 In total, eleven different HPV genotyping assays from 8 different manufacturers will be evaluated in the
149 VALGENT4 study (Table 1), using GP5+/6+ PCR-EIA with genotyping as comparator in line with previous
150 VALGENT installments.

151 **Sample collection**

152 The VALGENT4 study was embedded into the routine cytology-based operations of the Department of
153 Pathology, Copenhagen University Hospital, Hvidovre, Denmark. This pathology department receives
154 virtually all cervical cytology screening samples from women residing in the Danish Capital Region, covering
155 almost a third of the entire Danish population. Women are invited for screening every three years at ages
156 23-49 and every fifth year at ages 50-59. An HPV based “check out test” is offered to women 60-64 years of
157 age prior to exiting the screening program at age 65. The Danish cervical cancer screening program is free of
158 charge at point of delivery. Samples are collected predominately by general practitioners or in some cases by
159 gynecologists. The Combi-brush (Rovers, Oss, Netherlands) is used for collection of cell material and all
160 cytology samples are collected in SurePath medium.

161 Samples from women aged 30-59 years were included in VALGENT4 study.

162 The VALGENT4 panel consists of 1,297 routine samples collected in two groups as previously described by
163 the VALGENT framework [9]:

- 164 - 998 unselected, consecutive samples, of which 51 samples had abnormal cytology (\geq atypical
165 squamous cells of undetermined origin, (ASCUS)). By subsequent histological follow-up, 13 women
166 were diagnosed with cervical intraepithelial neoplasia 2 or more (\geq CIN2)
- 167 - 299 consecutive routine samples enriched for \geq ASCUS cytology (100 ASCUS, 100 low-grade
168 squamous intraepithelial lesions (LSIL), 99 high-grade squamous intraepithelial lesions (HSIL)). From
169 this population, 106 women were presented with \geq CIN2 upon follow up.

170 Three samples were discarded after collection due to lack of general consent after mandatory cross-check in
171 *Danish human biological material in health research projects register (Vævsanvendelsesregisteret).*

172 **Sample processing**

173 SurePath samples nominally contain 10 ml of liquid upon arrival in the laboratory. Currently, the Department
174 of Pathology processes cervical cytology samples using the BD Totalys instrumentation. Here, the cytology
175 processing uses 8 ml of the original sample, which is column purified for debris and mucus, spun down,
176 pelleted, and re-suspended in 1 ml BD Density Reagent. Of this 1 ml medium, 200 µl is used for the cytology
177 slide procedure. In routine, the remaining 800 µl is automatically added the 2 ml of fresh SurePath medium,
178 resulting in a residual volume of 2.8 ml. By calculation, the final volume of the pelleted material in the 2.8 ml
179 holds a 2.3 times higher concentration of cellular material than the original material. In order to approximate
180 the cellularity of the original 10 ml sample, we developed a reconstitution protocol utilizing post-cytology
181 pelleted material mixed with residual original material. The post-cytology pelleted sample was reconstituted
182 by adding 3.6 ml of fresh SurePath medium, resulting in a volume of 6.4 ml (Figure 1). In addition, the residual,
183 original 2 ml surplus material from the cytology procedure was added to the 6.4 ml of reconstituted sample.
184 Consequently, the nominal volume of the reconstituted sample available for testing in the VALGENT4
185 protocol was 8.4 ml.

186 The reconstituted samples were split into aliquots for DNA extraction and original sample material upon
187 processing of the samples (see processing time above).

188 **DNA extraction**

189 DNA was extracted under strictly controlled conditions using standard operating procedure clinically routine
190 DNA extraction protocol at the parent laboratory based upon the MagNA Pure96 system (Roche Diagnostics,
191 Pleasanton, CA). In short, 4 x1 ml aliquots of reconstituted material were transferred to four Eppendorf tubes,
192 spun down (14,000 rpm, 5 min), the supernatant removed, and the pellet re-suspended in a mix of 180 µl
193 phosphate buffered saline (10x conc. pH 7.4, Pharmacy product) and 20 µl Proteinase K (Recombinant, PCR
194 grade, Roche Diagnostics). The samples were vortexed and incubated one hour at 56°C for Proteinase K
195 digestion. Subsequently, the tubes were incubated for one hour at 90°C to reverse SurePath formaldehyde

196 induced co-valent cross linking. The entire volume was transferred to the MagNA Pure 96 system and
197 extracted using the MagNA Pure LC Total Nucleic Acid Isolation Kit (Roche Diagnostics). The four DNA eluates
198 were pooled into one tube to a total volume of 400 µl.

199 **Cytology**

200 Cytology was reported according to the Bethesda 2001 criteria, using computer assisted reading by BD Focal
201 Point and Slide Wizard systems. ASCUS cases were triaged (reflex testing) routinely with HPV testing (BD
202 Onclarity). All abnormal cytology findings were adjudicated as per standard operating procedures of the
203 parent laboratory by a senior pathologist. Women with LSIL were recommended for repeated cytology
204 testing after 6 months. Women with inadequate cytology were recommended for repeated sampling and
205 testing within 3 months. Women with normal cytology were returned to routine screening according to age-
206 specific intervals. The complete screening history of the women included in VALGENT4 was retrieved from
207 the Danish Pathology Data Bank (PatoBank). All clinical follow up was managed according to Danish
208 guidelines, and the outcome of the VALGENT4 HPV testing did not affect clinical follow up recommendations.
209 All HPV testing performed in the framework of VALGENT4 was blinded to cytology and clinical follow-up.

210 **Histology**

211 All histology available for disease ascertainment was derived from the clinical follow-up of women whose
212 samples were included in the VALGENT4 cohort. The following screening sample outcomes elicit a
213 recommendation for follow-up with colposcopy and biopsy-taking under the current Danish guidelines:

- 214 • Women, ≥ 30 years with ASCUS and HPV-positivity upon triage testing.
- 215 • Women with HSIL, atypical squamous cells – cannot exclude HSIL (ASC-H), atypical glandular cells
216 (AGS) or cytological indications of carcinoma.
- 217 • Women with continued ASCUS and LSIL cytological diagnosis, as evaluated by the present sample
218 and screening history.

219 Danish Gynecology Guidelines recommend biopsy taking on all aceto-white lesions observed, or a random
220 four quadrants biopsy set where no lesions are visible upon colposcopy. The histological follow-up for all
221 women included in the VALGENT4 study was retrieved from the PatoBank.

222 **Data source for clinical outcomes**

223 The PatoBank follow-up period was on average 19 months (min. 18 months; max 20 months) from collection
224 of the VALGENT4 samples. The registration of cervical cytology and histology in the PatoBank has been
225 complete on a national level since 2008.

226 **Participating Scientific Partner panel testing**

227 All HPV testing within VALGENT4 was blinded to cytology and histology outcomes from the parent laboratory
228 as well as the HPV test results from other scientific partners. All test results were reported to the Scientific
229 Institute of Public Health, Brussels, where the data are compiled, and performance calculations are
230 conducted [9]. All HPV testing was performed according to manufacturer's specifications. Nine different
231 participating scientific partners (including the parent site) are taking part in the VALGENT4 study (Table 1).

232 **Pre-analytical characterization of LBC material and DNA aliquot panel samples.**

233 Pre-analytical characterization of cellularity of reconstituted original LBC material

234 All 1,297 panel samples were tested with the BD Onclarity HPV assay (Onclarity) on the automated Viper Lt
235 platform under the VALGENT4 protocol. The Onclarity assay harbors an internal Human Beta globin (HBB)
236 control for assay performance and sample adequacy. To evaluate the resulting cellularity of the resuspension
237 protocol, mean HBB Ct values from the 1,297 VALGENT4 samples were retrieved from the Viper Lt and used
238 as a proxy marker for cellularity. As comparator, mean HBB Ct values of unselected, unannotated routine
239 cervical screening SurePath samples were used as a reference group for cellularity (designated "Control-1",
240 N=293). The Onclarity assay on the Viper Lt Platform workflow has previously been described in detail [27,
241 36]. In summary, 0.5 ml original SurePath material is transferred to a BD CBD medium tube prior to heat
242 treatment for 30 min at 120°C on the BD pre-warm station. The pre-warmed samples are subsequently

243 transferred to the fully automated Viper Lt platform and tested with the Onclarity assay according to
244 manufacturer's recommendations. Mean HBB values between the two groups were compared using one-
245 way ANOVA testing (IBM SPSS ver. 22).

246 Amplifiable DNA copies and DNA fragmentation evaluation of DNA extracts

247 The iPLEX PRO ExomeQC assay (Agena Bioscience, Hamburg) is a quantitative assay which reports average
248 amplifiable human DNA copies in a sample, in addition to reporting the relative number of amplifiable
249 fragments of 100, 200, 300, 400, and 500 base pairs (bp) lengths. Here, the assay was used as a quality control
250 of the DNA extracts derived from the reconstituted VALGENT4 panel. As comparator, DNA from unselected,
251 unannotated, and un-manipulated SurePath cervical screening samples collected from the routine service
252 were used as reference (designated "Control-2", N=184).

253 The ExomeQC panel harbors 44 Single Nucleotide Polymorphisms (SNPs), three markers for gender
254 identification and five copy-number markers in a single multiplexed assay. The assay uses the Matrix Assisted
255 Laser Desorption/Ionization-Time of Flight Mass Spectrometry (MALDI-TOF MS) technology. An initial
256 multiplex PCR amplification was set up with 2 µl DNA (duration: 2.5 hours), followed by a Shrimp Alkaline
257 Phosphatase (SAP) reaction (dephosphorylate excess Nucleotides, duration: 50 min). This was followed by an
258 iPLEX Pro single base extension PCR reaction in which a mix of oligonucleotide extension primers designed
259 to anneal to the amplified DNA fragments, was added together with an extension enzyme and mass-modified
260 dideoxynucleotide terminators (duration 3.5 hours). The extension products were subsequently desalted
261 with Clean Resin prior to being loaded into the MassArray Dx Nanodispenser RS1000 (Agena Bioscience,
262 Hamburg), which transfers the analyte to a spectroCHIP. Here the sample crystalize with the matrix on the
263 chip, which was then analyzed on the MassARRAY Dx Analyzer 4 (Agena Bioscience, Hamburg). The crystals
264 were irradiated by a laser, inducing desorption and ionization. The MA4 accelerates the samples to a detector
265 that differentiates genetic variants by molecular mass.

266 **Statistical analysis**

267 The Onclarity assay has a three well set-up with nine HPV genotype read-outs with internal HBB control
268 included in each well. The Ct-value of the HBB was calculated as an average of the three HBB Ct values in
269 each of the three wells for every panel and control sample. The Onclarity cut-off for all channels is Ct 34.2.
270 For the ExomeQC analysis, inadequate samples (insufficient number of SNPs detected) and samples with
271 outlier values above 10,000,000 units were excluded in the resulting analysis. Mean Ct, mean available
272 copy numbers, 95% confidence interval, standard deviations as well as the one-way ANOVA test used to
273 calculate the statistical difference in cellularity and DNA integrity parameters between the VALGENT4 panel
274 and Control-1 and 2 samples were calculated using IBM SPSS statistics 22.

275 **Results**

276 The overall characteristics of the VALGENT4 study population are summarized in Table 2. The mean age
277 was 42.8 years for all included women, and 40.4 and 42.2 years for the unselected, consecutive and
278 enriched subpopulations, respectively. In the 998 consecutive, unselected panel samples, 5.1% (51
279 samples) had abnormal cytology and as of March 2018, 1.3% (13 samples) had \geq CIN2 follow-up histology
280 registered in the Danish National Pathology Registry (PatoBank). The enriched population of 100 ASCUS,
281 100 LSIL and 99 HSIL cytology samples resulted in 106 \geq CIN2 histology samples as per register update
282 March 2018. In total, 119 \geq CIN2 cases were registered at data retrieval \geq 20 month after baseline.

283 Comparing the relative cellularity (mean Ct values of HBB control from the Onclarity assay) on the 1,297
284 LBC panel samples with data from 293 separately collected, un-manipulated routine cytology screening
285 samples (Denominated "Control-1" samples, Table 3A). The mean Ct value of the VALGENT4 samples
286 (Ct=24.8) was found to be on average 1.0 Ct above the mean Ct value of the Control-1 samples (Ct=23.8),
287 indicating slightly less cellularity in the panel samples compared to native samples.

288 For DNA extracted panel samples, characterization was done using the Exome iPLEX pro assay, which
289 reports amplifiable copies on individual samples as well as amplifiable copies at five different base pair

290 lengths (Table 3B). Here, the data showed statistically difference in the number of amplifiable DNA copies
291 comparing the VALGENT4 panel to the Control-2 population (ratio of 0.7, p-value<0.01). When looking at
292 the relative proportion of different base pair length fragments, the statistical difference was only evident at
293 the longest 400 and 500 bp fragments (Table 3B).

294 **Discussion**

295 The VALGENT framework is an international cooperation aimed at evaluating HPV genotyping tests for
296 clinical use in cervical cancer screening. Here we present the study protocol for the fourth installment of
297 the VALGENT framework, presenting data that provides insight into the underlying quality of the 1,297
298 VALGENT4 samples. One of the novel elements in VALGENT4 is the resuspension protocol employed which
299 attempts to emulate the original cellularity of un-manipulated screening samples (Figure 1).

300 For research use, most cross-sectional studies where HPV tests are applied rely on stored, residual clinical
301 cytology screening material. In contrast, most HPV tests are intended to run on recently collected and un-
302 manipulated original sample material. In addition, for evaluations where multiple assays are applied to
303 samples, the volume of sample may be a rate-limiter. Consequently, to reconcile the need for authenticity
304 with the operational requirements of the study, we developed a reconstitution protocol taking into account
305 the changes introduced to the freshly collected VALGENT4 samples by routine cytology processing. The aim
306 was to reestablish the residual cytology pelleted material to an approximate *per volume cellularity*
307 corresponding to the original cellularity in un-manipulated clinical samples (Figure 1). This reconstitution
308 protocol was developed with the assistance of Becton, Dickinson and Company, the manufacturer of the
309 SurePath medium, by reverse engineering the concentration and resuspension steps included in standard
310 SurePath processing of original, fresh samples.

311 As the mean beta globin Ct value of the VALGENT4 samples (Ct=24.8) was found to be 1.0 Ct above the
312 Control-1 samples (Ct=23.8), this suggests that the reconstituted panel samples had slightly lower

313 cellularity than freshly collected, un-manipulated screening samples. However, in the context of the clinical
314 positive-negative Onclarity HPV assay cut off (Ct 34.2), the reconstituted panel samples were fully
315 acceptable as clinically, analytical material for HPV analysis. In combination with the ratio data from the
316 ExomeQC analysis, this describes the resulting degree of dilution introduced to the reconstituted panel
317 samples compared to the original sample cellularity. Moreover, the DNA fragmentation analysis showed
318 that fresh, un-manipulated samples had more 400 and 500 bp DNA fragments compared to the
319 reconstituted panel samples. This clear trend downwards between reconstituted samples and original
320 samples in Control-2 was however not unexpected as all sorts of manipulation of cells and DNA leads to
321 some level of deterioration. Whether this has implications for the analytical and clinical performance of
322 HPV assays is unlikely, given that all included assays in the VALGENT4 protocol rely on amplification of less
323 than 500bp HPV fragments (Table 1).

324 In conclusion, the two QA/QC methods employed indicate that the reconstituted samples used for the
325 VALGENT evaluation contain slightly less analytical material compared to fresh, un-manipulated screening
326 samples. However, the amount of analytical relevant material is reassuringly within the operational
327 boundaries of all included VALGENT4 HPV assays (Table 1).

328 Besides the reconstitution protocol, VALGENT4 introduced central organization of DNA extraction for all
329 assays running on this input material. This was chosen to eliminate any variation introduced at the level of
330 DNA extraction methodology. Of the 11 assays evaluated in VALGENT4, only BD Onclarity, Seegene Anyplex
331 HPV28/HPV II HR and Roche Cobas HPV assays required original SurePath LBC material as input material for
332 analysis.

333 Until now, three VALGENT installments have been completed with several genotyping assays tested. The data
334 from these VALGENT studies has showed that the VALGENT framework provides a good base for cross-
335 sectional clinical validation of HPV genotyping assays, by use of well annotated cervical screening samples
336 collected in various screening programs around Europe [9]. A common feature of the VALGENT installments

337 is the use of GP5+/6+ PCR-EIA as the comparator assay to all other included assays. This brings the VALGENT
338 in line with the stipulations within the international validation Guidelines of Meijer et al.

339 The first VALGENT study, VALGENT-1, included cervical cancer screening samples provided by the AML
340 laboratory (Antwerp, Belgium) and included validation of four full genotyping assays [37-39]. The VALGENT-
341 1 framework included SurePath samples from biobanked screening samples, and thus not freshly collected
342 samples as included here. The second VALGENT study, VALGENT-2, used ThinPrep, involved the collection
343 of samples from the Scottish Cervical Screening Programme via the Scottish HPV Reference laboratory and
344 involved the evaluation of four assays with genotyping capability [40-43]. VALGENT-3 also included
345 ThinPrep collected cervical cancer screening samples, this time from the Laboratory for Molecular
346 Microbiology, Faculty of Medicine, University of Ljubljana, Slovenia [44] and evaluated 10 different HPV
347 assays with different degrees of genotyping capability. VALGENT-1 [37-39] and VALGENT-2 [40-43] have
348 both provided several peer-reviewed papers and publications from VALGENT-3 are underway [32, 44].
349 Formal clinical validation of HPV assays for use in screening has primarily been undertaken on ThinPrep
350 collected samples [24, 26, 28, 30, 41, 45, 46], with only one assay to date, the BD Onclarity, being validated
351 on both ThinPrep [26] *and* SurePath collected samples [27]. Following this, it will be interesting to see how
352 the HPV genotyping assays included in VALGENT4 perform on SurePath collected screening samples in this
353 large-scale performance comparison.

354 **Ethical and Data Inspection Agency approvals:**

355 The study was approved by the Danish Data Inspection Agency J. No. AHH-2017-024, I-Suite: 05356. EU-
356 GDPR compliant data handler agreement was established between the principal site Hvidovre Hospital and
357 the Statistical analysis unit at Sciensano, Brussels. All collected samples were verified for non-compliance in
358 the Danish human biological material in health research projects register (Vævsanvendelsesregisteret).

359

360 **Acknowledgement**

361 We wish to thank the staff of the Molecular Pathology Laboratory, Dept. Pathology, Hvidovre Hospital,
362 Copenhagen, for their diligent work on assembling, aliquoting and processing the Valgent4 panel samples.
363 Moreover, we would like to thank all the participating company partners who engaged in this private-public
364 research project.

365 **Declaration of competing interest**

366 VALGENT is an independent researcher induced research project, set up by Sciensano, where
367 manufacturers can have their HPV assays evaluated, under condition to provide equipment and kits and to
368 cover costs for laboratory work and statistical analysis. Manufacturers cannot influence publication of
369 manuscripts.

370 Jesper Bonde attended meetings with various HPV device manufacturers. JB has received honoraria from
371 Hologic/Gen-probe, Roche, Qiagen, Genomica, and BD Diagnostics for lectures. Hvidovre Hospital has
372 ongoing contracts with BD Diagnostics, Genomica, Self-Screen and EU Horizon2020.

373 Ditte Møller Ejegod attended meetings with various HPV device manufacturers.

374 Joakim Dillner has previously received research grants to his institution for research on HPV vacciens from
375 an HPV vaccine manufacturer (Merck/SPMSD). The Karolinska University Hospital has ongoing contracts
376 with several manufacturers of HPV diagnostics, including Roche and Genomica.

377 Daniëlle Heideman: minority stock portfolio in Self-screen B.V., a spin-off company of VU University
378 Medical Center Amsterdam. Self-screen B.V holds patents related to the work, and has developed and
379 manufactured the HPV-Risk Assay. DAMH has been on the speaker's bureau of Qiagen and serves
380 occasionally on the scientific advisory board of Pfizer and Bristol-Meyer Squibb.

381 Wim Quint: No conflicts of interest to declare

382 Miguel Angel Pavon Ribas: received reagents for HPV testing at no cost from Roche, SeeGene, Qiagen and
383 Genomica for research purposes.

384 Elizaveta Padalko: No conflicts of interest to declare

385 Irene Kraus Christiansen: No conflicts of interest to declare

386 Kate Cuschieri: No conflicts of interest to declare, KCs institution has received research funding and or
387 associated gratis consumable from the following in the last 3 years: Hologic, Cepheid, Qiagen, Becton-
388 Dickinson, Euroimmun, SelfScreen, LifeRiver, Genomica, Genefirst

389 Marc Arbyn and Lan Xu were supported by COHEAHR Network (grant No. 603019), coordinated by the Free
390 University of Amsterdam (The Netherlands), funded by the 7th Framework Programme of DG Research and
391 Innovation, European Commission (Brussels, Belgium)

392 Copenhagen University Hospital, Hvidovre: Reagents and instrumentation for testing Onclarity, CLART
393 HPV4s and MassArray HPV test were received free of charge from the manufacturers for the duration of
394 the testing. Limited co-funding for the project was obtained as part of the collaboration agreement with
395 the manufacturers Becton, Dickinson and Company, Genomica SAU, and Agena Biotech GmbH.

396 Norwegian HPV Reference Lab: No external funding was obtained for this study

397 Ghent University: Reagents for testing INNO-LIPA genotyping Extra II test was received free of charge from
398 the manufacturer for the duration of the testing.

399 Infection and Cancer Laboratory, Barcelona: No external funding was obtained for this study

400 VU University Medical center: No external funding was obtained for this study

401 Scottish HPV reference lab: Reagents and instrumentation for testing LifeRiver Venus, LifeRiver Harmonia,
402 and Genefirst HPV-MPA Genotyping Test were received free of charge from the manufacturers for the

403 duration of the testing. Limited co-funding for the project was obtained as part of the collaboration
404 agreement with the manufacturers Zhanghai Biotech (LifeRiver) and GeneFirst (HPV-MPA Genotyping Test)
405 DDL Diagnostic Laboratory: No external funding was obtained for this study
406 Karolinska University hospital: No external funding was obtained for this study.
407 Sciensano, the institute where M.A. is employed, received support from Becton Dickinson, Genomica SAU,
408 Agena Biotech GmbH, Zhanghai Biotech (LifeRiver), GeneFirst and FujiRebio for methodological and
409 statistical work as described in the VALGENT Network (Arbyn et al, J Clin Virol 2016).

410 References

- 411 1. Bouvard V, Baan R, Straif K, Grosse Y, Secretan B, El Ghissassi F, Benbrahim-Tallaa L, Guha N, Freeman
412 C, Galichet L *et al*: **A review of human carcinogens--Part B: biological agents**. *Lancet Oncol* 2009,
413 **10(4):321-322**.
- 414 2. Group IMw: **IARCH Monographs on the Evaluation of Carconogenic Risks of Humans**. *Human*
415 *Papillomavirus, Lyon: IARCPress* 2007, **90**.
- 416 3. Petry KU, Liebrich C, Luyten A, Zander M, Iftner T: **Surgical staging identified false HPV-negative**
417 **cases in a large series of invasive cervical cancers**. *Papillomavirus Res* 2017, **4:85-89**.
- 418 4. Cubie HA, Cuschieri K: **Understanding HPV tests and their appropriate applications**. *Cytopathology*
419 *: official journal of the British Society for Clinical Cytology* 2013, **24(5):289-308**.
- 420 5. Arbyn M, Ronco G, Anttila A, Meijer CJ, Poljak M, Ogilvie G, Koliopoulos G, Naucler P,
421 Sankaranarayanan R, Peto J: **Evidence regarding human papillomavirus testing in secondary**
422 **prevention of cervical cancer**. *Vaccine* 2012, **30 Suppl 5:F88-99**.
- 423 6. Arbyn M, Roelens J, Simoens C, Buntinx F, Paraskevaidis E, Martin-Hirsch PP, Prendiville WJ: **Human**
424 **papillomavirus testing versus repeat cytology for triage of minor cytological cervical lesions**.
425 *Cochrane Database Syst Rev* 2013(3):CD008054.
- 426 7. Ronco G, Dillner J, Elfstrom KM, Tunesi S, Snijders PJ, Arbyn M, Kitchener H, Segnan N, Gilham C,
427 Giorgi-Rossi P *et al*: **Efficacy of HPV-based screening for prevention of invasive cervical cancer:**
428 **follow-up of four European randomised controlled trials**. *Lancet* 2014, **383(9916):524-532**.
- 429 8. Poljak M, Cuzick J, Kocjan BJ, Iftner T, Dillner J, Arbyn M: **Nucleic acid tests for the detection of alpha**
430 **human papillomaviruses**. *Vaccine* 2012, **30 Suppl 5:F100-106**.
- 431 9. Arbyn M, Depuydt C, Benoy I, Bogers J, Cuschieri K, Schmitt M, Pawlita M, Geraets D, Heard I, Gheit
432 T *et al*: **VALGENT: A protocol for clinical validation of human papillomavirus assays**. *J Clin Virol* 2016,
433 **76 Suppl 1:S14-21**.
- 434 10. Khan MJ, Castle PE, Lorincz AT, Wacholder S, Sherman M, Scott DR, Rush BB, Glass AG, Schiffman M:
435 **The elevated 10-year risk of cervical precancer and cancer in women with human papillomavirus**
436 **(HPV) type 16 or 18 and the possible utility of type-specific HPV testing in clinical practice**. *J Natl*
437 *Cancer Inst* 2005, **97(14):1072-1079**.
- 438 11. Castellsague X, Ault KA, Bosch FX, Brown D, Cuzick J, Ferris DG, Joura EA, Garland SM, Giuliano AR,
439 Hernandez-Avila M *et al*: **Human papillomavirus detection in cervical neoplasia attributed to 12**
440 **high-risk human papillomavirus genotypes by region**. *Papillomavirus Res* 2016, **2:61-69**.

- 441 12. Cuzick J, Wheeler C: **Need for expanded HPV genotyping for cervical screening.** *Papillomavirus Res*
442 2016, **2**:112-115.
- 443 13. Joura EA, Ault KA, Bosch FX, Brown D, Cuzick J, Ferris D, Garland SM, Giuliano AR, Hernandez-Avila
444 M, Huh W *et al*: **Attribution of 12 high-risk human papillomavirus genotypes to infection and**
445 **cervical disease.** *Cancer Epidemiol Biomarkers Prev* 2014, **23**(10):1997-2008.
- 446 14. Smelov V, Elfstrom KM, Johansson AL, Eklund C, Naucler P, Arnheim-Dahlstrom L, Dillner J: **Long-term**
447 **HPV type-specific risks of high-grade cervical intraepithelial lesions: a 14-year follow-up of a**
448 **randomized primary HPV screening trial.** *Int J Cancer* 2015, **136**(5):1171-1180.
- 449 15. Kjaer SK, Munk C, Junge J, Iftner T: **Carcinogenic HPV prevalence and age-specific type distribution**
450 **in 40,382 women with normal cervical cytology, ASCUS/LSIL, HSIL, or cervical cancer: what is the**
451 **potential for prevention?** *Cancer Causes Control* 2014, **25**(2):179-189.
- 452 16. Arbyn M, Tommasino M, Depuydt C, Dillner J: **Are 20 human papillomavirus types causing cervical**
453 **cancer?** *J Pathol* 2014, **234**(4):431-435.
- 454 17. Bzhalava D, Guan P, Franceschi S, Dillner J, Clifford G: **A systematic review of the prevalence of**
455 **mucosal and cutaneous human papillomavirus types.** *Virology* 2013, **445**(1-2):224-231.
- 456 18. Li N, Franceschi S, Howell-Jones R, Snijders PJ, Clifford GM: **Human papillomavirus type distribution**
457 **in 30,848 invasive cervical cancers worldwide: Variation by geographical region, histological type**
458 **and year of publication.** *Int J Cancer* 2011, **128**(4):927-935.
- 459 19. Kjaer SK, Frederiksen K, Munk C, Iftner T: **Long-term absolute risk of cervical intraepithelial**
460 **neoplasia grade 3 or worse following human papillomavirus infection: role of persistence.** *J Natl*
461 *Cancer Inst* 2010, **102**(19):1478-1488.
- 462 20. Serrano B, de Sanjose S, Tous S, Quiros B, Munoz N, Bosch X, Alemany L: **Human papillomavirus**
463 **genotype attribution for HPVs 6, 11, 16, 18, 31, 33, 45, 52 and 58 in female anogenital lesions.** *Eur*
464 *J Cancer* 2015, **51**(13):1732-1741.
- 465 21. Kavanagh K, Pollock KG, Cuschieri K, Palmer T, Cameron RL, Watt C, Bhatia R, Moore C, Cubie H,
466 Cruickshank M *et al*: **Changes in the prevalence of human papillomavirus following a national**
467 **bivalent human papillomavirus vaccination programme in Scotland: a 7-year cross-sectional study.**
468 *Lancet Infect Dis* 2017, **17**(12):1293-1302.
- 469 22. Huh WK, Ault KA, Chelmow D, Davey DD, Goulart RA, Garcia FA, Kinney WK, Massad LS, Mayeaux EJ,
470 Saslow D *et al*: **Use of primary high-risk human papillomavirus testing for cervical cancer screening:**
471 **interim clinical guidance.** *Obstet Gynecol* 2015, **125**(2):330-337.
- 472 23. Arbyn M, Xu L, Verdoodt F, Cuzick J, Szarewski A, Belinson JL, Wentzensen N, Gage JC, Khan MJ:
473 **Genotyping for Human Papillomavirus Types 16 and 18 in Women With Minor Cervical Lesions: A**
474 **Systematic Review and Meta-analysis.** *Ann Intern Med* 2017, **166**(2):118-127.
- 475 24. Meijer CJ, Berkhof J, Castle PE, Hesselink AT, Franco EL, Ronco G, Arbyn M, Bosch FX, Cuzick J, Dillner
476 J *et al*: **Guidelines for human papillomavirus DNA test requirements for primary cervical cancer**
477 **screening in women 30 years and older.** *International journal of cancer Journal international du*
478 *cancer* 2009, **124**(3):516-520.
- 479 25. Arbyn M, Snijders PJ, Meijer CJ, Berkhof J, Cuschieri K, Kocjan BJ, Poljak M: **Which high-risk HPV**
480 **assays fulfil criteria for use in primary cervical cancer screening?** *Clin Microbiol Infect* 2015,
481 **21**(9):817-826.
- 482 26. Ejegod D, Serrano, I., Cuschieri, KS., Nussbaumer, WA., Vaughan, LM., Ahmad, AS., Cuzick, J., Bonde.,
483 J.: **Clinical Validation of the BD Onclarity HPV Assay Using a Non-Inferiority Test.** *Medical*
484 *Microbiology & Diagnosis* 2013, **S3**(doi:10.4172/2161-0703.S3-003).
- 485 27. Ejegod D, Bottari F, Pedersen H, Sandri MT, Bonde J: **The BD Onclarity HPV Assay on Samples**
486 **Collected in SurePath Medium Meets the International Guidelines for Human Papillomavirus Test**
487 **Requirements for Cervical Screening.** *Journal of clinical microbiology* 2016, **54**(9):2267-2272.
- 488 28. Heideman DA, Hesselink AT, Berkhof J, van Kemenade F, Melchers WJ, Daalmeijer NF, Verkuijten M,
489 Meijer CJ, Snijders PJ: **Clinical validation of the cobas 4800 HPV test for cervical screening purposes.**
490 *J Clin Microbiol* 2011, **49**(11):3983-3985.

- 491 29. Hesselink AT, Heideman DA, Berkhof J, Topal F, Pol RP, Meijer CJ, Snijders PJ: **Comparison of the**
492 **clinical performance of PapilloCheck human papillomavirus detection with that of the GP5+/6+-**
493 **PCR-enzyme immunoassay in population-based cervical screening.** *Journal of clinical microbiology*
494 2010, **48**(3):797-801.
- 495 30. Hesselink AT, Meijer CJ, Poljak M, Berkhof J, van Kemenade FJ, van der Salm ML, Bogaarts M, Snijders
496 PJ, Heideman DA: **Clinical validation of the Abbott RealTime High Risk HPV assay according to the**
497 **guidelines for human papillomavirus DNA test requirements for cervical screening.** *J Clin Microbiol*
498 2013, **51**(7):2409-2410.
- 499 31. Hesselink AT, Sahli R, Berkhof J, Snijders PJ, van der Salm ML, Agard D, Bleeker MC, Heideman DA:
500 **Clinical validation of Anyplex II HPV HR Detection according to the guidelines for HPV test**
501 **requirements for cervical cancer screening.** *J Clin Virol* 2016, **76**:36-39.
- 502 32. Xu L, Ostrbenk A, Poljak M, Arbyn M: **Assessment of the Roche Linear Array HPV Genotyping Test**
503 **within the VALGENT framework.** *J Clin Virol* 2018, **98**:37-42.
- 504 33. Eklund C, Forslund O, Wallin KL, Dillner J: **Continuing global improvement in human papillomavirus**
505 **DNA genotyping services: The 2013 and 2014 HPV LabNet international proficiency studies.** *J Clin*
506 *Virol* 2018, **101**:74-85.
- 507 34. Muhr LSA, Eklund C, Dillner J: **Towards quality and order in human papillomavirus research.** *Virology*
508 2018, **519**:74-76.
- 509 35. Pedersen K, Fogelberg S, Thamsborg LH, Clements M, Nygard M, Kristiansen IS, Lynge E, Sparen P,
510 Kim JJ, Burger EA: **An overview of cervical cancer epidemiology and prevention in Scandinavia.** *Acta*
511 *Obstet Gynecol Scand* 2018, **97**(7):795-807.
- 512 36. Ejegod DM, Junge J, Franzmann M, Kirschner B, Bottari F, Sideri M, Sandri MT, Bonde J: **Clinical and**
513 **analytical performance of the BD Onclarity HPV assay for detection of CIN2+ lesions on SurePath**
514 **samples.** *Papillomavirus Res* 2016, **2**:31-37.
- 515 37. Schmitt M, Depuydt C, Benoy I, Bogers J, Antoine J, Pawlita M, Arbyn M, group Vs: **Viral load of high-**
516 **risk human papillomaviruses as reliable clinical predictor for the presence of cervical lesions.**
517 *Cancer Epidemiol Biomarkers Prev* 2013, **22**(3):406-414.
- 518 38. Schmitt M, Depuydt C, Benoy I, Bogers J, Antoine J, Arbyn M, Pawlita M, Group VS: **Prevalence and**
519 **viral load of 51 genital human papillomavirus types and three subtypes.** *Int J Cancer* 2013,
520 **132**(10):2395-2403.
- 521 39. Schmitt M, Depuydt C, Benoy I, Bogers J, Antoine J, Arbyn M, Pawlita M, Group VS: **Multiple human**
522 **papillomavirus infections with high viral loads are associated with cervical lesions but do not**
523 **differentiate grades of cervical abnormalities.** *J Clin Microbiol* 2013, **51**(5):1458-1464.
- 524 40. Cuschieri K, Geraets D, Cuzick J, Cadman L, Moore C, Vanden Broeck D, Padalko E, Quint W, Arbyn M:
525 **Performance of a Cartridge-Based Assay for Detection of Clinically Significant Human**
526 **Papillomavirus (HPV) Infection: Lessons from VALGENT (Validation of HPV Genotyping Tests).** *J Clin*
527 *Microbiol* 2016, **54**(9):2337-2342.
- 528 41. Cuschieri K, Geraets DT, Moore C, Quint W, Duvall E, Arbyn M: **Clinical and Analytical Performance**
529 **of the Onclarity HPV Assay Using the VALGENT Framework.** *Journal of clinical microbiology* 2015,
530 **53**(10):3272-3279.
- 531 42. Geraets DT, Cuschieri K, de Koning MN, van Doorn LJ, Snijders PJ, Meijer CJ, Quint WG, Arbyn M:
532 **Clinical evaluation of a GP5+/6+-based luminex assay having full high-risk human papillomavirus**
533 **genotyping capability and an internal control.** *J Clin Microbiol* 2014, **52**(11):3996-4002.
- 534 43. Heard I, Cuschieri K, Geraets DT, Quint W, Arbyn M: **Clinical and analytical performance of the**
535 **PapilloCheck HPV-Screening assay using the VALGENT framework.** *J Clin Virol* 2016, **81**:6-11.
- 536 44. Polman NJ, Ostrbenk A, Xu L, Snijders PJF, Meijer C, Poljak M, Heideman DAM, Arbyn M: **Evaluation**
537 **of the clinical performance of the HPV-Risk assay using the VALGENT-3 panel.** *J Clin Microbiol* 2017.
- 538 45. Boers A, Wang R, Slagter-Menkema L, van Hemel BM, Ghysaert H, van der Zee AG, Wisman GB,
539 Schuurin E: **Clinical validation of the Cervista HPV HR test according to the international guidelines**

- 540 **for human papillomavirus test requirements for cervical cancer screening. *J Clin Microbiol* 2014,**
541 **52(12):4391-4393.**
- 542 46. Heideman DA, Hesselink AT, van Kemenade FJ, Iftner T, Berkhof J, Topal F, Agard D, Meijer CJ, Snijders
543 PJ: **The Aptima HPV assay fulfills the cross-sectional clinical and reproducibility criteria of**
544 **international guidelines for human papillomavirus test requirements for cervical screening. *J Clin***
545 ***Microbiol* 2013, 51(11):3653-3657.**
- 546
- 547

548

Valgent-4 included Assays	Aliquot	Amplicon length	Scientific partner
BD Onclarity HPV Assay	Original Material	79-137 bp.	Copenhagen University Hospital, Pathology Laboratory, Hvidovre, Denmark
Genomica CLART HPV4 assay	DNA	465 bp.	
Agena HPV MassArray assay	DNA	90-122 bp.	
Roche cobas 4800 HPV Test	Original Material	~200 bp	Norwegian HPV Reference Lab, Akerhus University Hospital Norway
Fujirebio INNO-LiPA Genotyping Extra II test	DNA	65 bp	Ghent University, Ghent, Belgium
SeeGene Anyplex HPV28 detection test	DNA or Original Material	~150 bp	Infection and Cancer Laboratory. Cancer Epidemiology Research Program, Barcelona, Spain
Seegene Anyplex II HPV test		~150 bp	
Self-screen HPV-Risk assay	DNA	~150 bp	VU University Medical Center, Amsterdam, The Netherlands
Genefirst HPV-MPA Genotyping Test	DNA	150 bp.	Scottish HPV Reference Lab, Royal Infirmary of Edinburgh Scotland
Liferiver Harmony test	DNA	100-200 bp.	
Liferiver Venus tes		100-200 bp.	
Comparator assays			
GP5+/6+ EIA Luminex	DNA		DDL Diagnostic Laboratory, Rijswijk, The Netherlands
GP5+/6+ PCR EIA kit HPV GP HR	DNA		International HPV Reference Center, Karolinska University Hospital, Stockholm, Sweden

549 **Table 1:** HPV genotyping assays evaluated, concurrent material required and scientific partners under the VALGENT4
550 study protocol

551

552

553

Characteristics	Screening population		Enriched population ¹		Total population	
	N	%	N	%	N	%
Age						
Mean Age in years	42.8 (36.0-49.0)		40.4 (34.0-45.8)		42.2 (36.0 to 48.0)	
30-39	383	38.4%	148	49.8%	531	41.0%
40-49	408	40.9%	111	37.4%	519	40.1%
50-59	207	20.7%	38	12.8%	245	18.9%
30-59	998	100%	297	100%	1,295	100%
Cytology						
NILM	947	94.9%	0	0%	947	73.0%
ASCUS	6	0.6%	100	33.3%	106	8.2%
LSIL	21	2.1%	100	33.3%	121	9.3%
HSIL. AGC. ASC-H	24	2.4%	99	33.3%	123	9.5%
Total	998	100%	299	100%	1,297	100%
Histology						
No histology	911	91.3%	55	18.4%	966	74.6%
Normal	57	5.7%	82	27.4%	139	10.7%
CIN1	17	1.7%	54	18.1%	71	5.5%
CIN2	5	0.5%	34	11.4%	39	3.0%
CIN3	7	0.7%	64	21.4%	71	5.5%
Cancer	1	0.1%	8	2.7%	9	0.7%
All histologies	87	8.7%	242	80.9%	329	25.4%
<CIN2	74	7.4%	136	45.5%	210	16.2%
≥CIN2	13	1.3%	106	35.5%	119	9.2%

554 **Table 2:** Overall characteristics of the VALGENT4 study population

555 1. Two women enrolled in Valgent4 each had two samples included; therefore therefor 299 samples from 297
556 women are included in the enriched cohort.

557

558

Group	N	Mean HBB Ct	Ratio ¹	St.d	95% Confidence Incidence		Min	Max	P-value ²
					Lower	Upper			
Valgent4	1,297	24.8	1.04	1.5	24.7	24.9	20.8	32.3	0.00
Control1	293	23.8	1.00	1.5	23.6	24.0	20.5	32.0	

559 **Table 3A:** Quality assurance of the VALGENT4 LBC samples: Mean HBB Ct value as approx. marker for cellularity

560 1: The Control1 population was used as a reference

561 2. p-value is calculated using the one-way ANOVA test

562

563

	Group	N ¹	Mean amplifiable copies	Ratio ²	Standard deviation	95% Confidence Incidence		Min	Max	P-value ³
						Lower	Upper			
Avg Copy	Valgent4	1,261	89,037	0.72	133,317	81,672	96,403	186	1,801,335	0.001
	Control2	182	123,942	1.00	114,870	107,141	140,743	3,000	879,343	
100 bp	Valgent4	1,261	34,782	1.06	84,948	30,089	39,475	369	1,633,069	0.757
	Control2	182	32,802	1.00	37,890	27,260	38,344	780	312,634	
200 bp	Valgent4	1,236	54,874	0.81	160,613	45,911	63,837	20	4,443,905	0.295
	Control2	181	67,766	1.00	104,680	52,413	83,119	569	795,228	
300 bp	Valgent4	1,261	156,201	0.77	380,175	135,198	177,205	193	8,254,550	0.124
	Control2	182	202,641	1.00	380,425	147,000	258,282	3,669	3,960,481	
400 bp	Valgent4	1,260	104,438	0.67	268,635	89,590	119,285	36	6,068,401	0.014
	Control2	182	156,585	1.00	268,209	117,356	195,813	3,726	2,904,381	
500 bp	Valgent4	1,249	93,544	0.59	396,191	71,550	115,537	48	8,947,189	0.031
	Control2	181	159,193	1.00	265,673	120,227	198,159	5,843	1,804,460	

564 **Table 3B:** Quality assurance of the VALGENT4 DNA samples: Amplifiable copy numbers at average and five different
565 base pair length

566 1: Samples with insufficient number of SNPs were excluded, in addition a sample could have a valid SNPs number, but
567 invalid result in one of the base pair groups, which mean that the numbers in the different base pair groups vary. In
568 addition, samples with outlier values of above 10,000,000 were excluded from both sample sets.

569 2: The Control2 population was used as a reference

570 3: The p-value is calculated using the one-way ANOVA test

571

572

573

